



Digital Receipt

This receipt acknowledges that **Turnitin** received your paper. Below you will find the receipt information regarding your submission.

The first page of your submissions is displayed below.

Submission author: Dewa Ayu Putu Sri Masyeni
Assignment title: Article
Submission title: Assessment of dengue and COVID-19 antibody rapid diagno...
File name: s12985-021-01522-2.pdf
File size: 783.43K
Page count: 5
Word count: 3,451
Character count: 18,661
Submission date: 14-Dec-2021 06:44PM (UTC+0700)
Submission ID: 1730139500

Santoso et al. *Viral J* (2021) 18:54
<https://doi.org/10.1186/s12985-021-01522-2>

Virology Journal

SHORT REPORT Open Access

Assessment of dengue and COVID-19 antibody rapid diagnostic tests cross-reactivity in Indonesia

Marsha S. Santoso¹, Sri Masyeni², Soetianingsih Haryanto², Benediktus Yohan¹, Martin L. Hibbert² and R. Tedjo Sasmono^{1*}

Abstract
Background: The coronavirus disease 2019 (COVID-19) pandemic remains ongoing around the world, including in areas where dengue is endemic. Dengue and COVID-19, to some extent, have similar clinical and laboratory features, which can lead to misdiagnosis, delayed treatment and patient's isolation. The use of rapid diagnostic tests (RDT) is easy and convenient for fast diagnosis, however there may be issues with cross-reactivity with antibodies for other pathogens.
Methods: We assessed the possibility of cross-reactivity between SARS-CoV-2 and dengue antibodies by (1) testing five brands of COVID-19 IgG / IgM RDTs on 60 RT-PCR confirmed dengue samples; (2) testing 95 RT-PCR confirmed COVID-19 samples on dengue RDT; and (3) testing samples positive for COVID-19 IgG and/or IgM on dengue RDT.
Results: We observed a high specificity across all five brands of COVID-19 RDTs, ranging from 98.3 to 100%. Out of the confirmed COVID-19 samples, one patient tested positive for dengue IgM only, another tested positive for dengue IgG only. One patient tested positive for dengue IgG, IgM, and NSI, suggesting a co-infection. In COVID-19 IgG and/or IgM samples, 6.3% of COVID-19 IgG-positive samples also tested positive for dengue IgG, while 21.1% of COVID-19 IgM-positive samples also tested positive for dengue IgG.
Conclusion: Despite the high specificity of the COVID-19 RDT, we observed cross-reactions and false-positive results between dengue and COVID-19. Dengue and COVID-19 co-infection was also found. Health practitioners in dengue endemic areas should be careful when using antibody RDT for the diagnosis of dengue during the COVID-19 pandemic to avoid misdiagnosis.
Keywords: COVID-19, RDT, IgG, IgM, Dengue, Specificity, Cross-reactivity

Background
Severe acute respiratory coronavirus 2 (SARS-CoV-2) emerged in Wuhan, China, causing a respiratory disease, coronavirus disease 2019 (COVID-19), and has now resulted in a global pandemic [1]. The pandemic remains ongoing in many countries, including areas where dengue is endemic, such as Indonesia, which adds a burden to health systems [2, 3]. There were over 130 000 reported cases of dengue in Indonesia in 2019 with an incidence rate of 51.48 cases per 100 000 population, an increase from the previous year's incidence of 24.75 cases per 100 000 population. As of 23 June, there are 68 000 cases of dengue reported across Indonesia in 2020, while COVID-19 cases continue to increase [4]. As of 26

*Correspondence: santoso@man.go.id
¹ Dengue Research Unit, Eijkman Institute for Molecular Biology, Ministry of Research and Technology/National Agency for Research and Innovation of the Republic of Indonesia, 4 Dikoreperan Rd, Jakarta 10430, Indonesia
Full list of author information is available at the end of the article

BMC

© The Author(s) 2021. **Open Access** This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons licence, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons licence, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons licence and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this licence, visit <http://creativecommons.org/licenses/by/4.0/>. The Creative Commons Public Domain Dedication waiver (<http://creativecommons.org/publicdomain/zero/1.0/>) applies to the data made available in this article, unless otherwise stated in a credit line to the data.

Assessment of dengue and COVID-19 antibody rapid diagnostic tests cross-reactivity in Indonesia

by Dewa Ayu Putu Sri Masyeni

Submission date: 14-Dec-2021 06:44PM (UTC+0700)

Submission ID: 1730139500

File name: s12985-021-01522-2.pdf (783.43K)

Word count: 3451

Character count: 18661

SHORT REPORT

Open Access



Assessment of dengue and COVID-19 antibody rapid diagnostic tests cross-reactivity in Indonesia

Marsha S. Santoso¹, Sri Masyeni², Sotianingsih Haryanto³, Benediktus Yohan¹, Martin L. Hibberd⁴ and R. Tedjo Sasmono^{1*} 

23

Abstract

Background: The coronavirus disease 2019 (COVID-19) pandemic remains ongoing around the world, including in areas where dengue is endemic. Dengue and COVID-19, to some extent, have similar clinical and laboratory features, which can lead to misdiagnosis, delayed treatment and patient's isolation. The use of rapid diagnostic tests (RDT) is easy and convenient for fast diagnosis, however there may be issues with cross-reactivity with antibodies for other pathogens.

Methods: We assessed the possibility of cross-reactivity between SARS-CoV-2 and dengue antibodies by: (1) testing five brands of COVID-19 IgG / IgM RDTs on 60 RT-PCR-confirmed dengue samples; (2) testing 95 RT-PCR-confirmed COVID-19 samples on dengue RDT; and (3) testing samples positive for COVID-19 IgG and/or IgM on dengue RDT.

Results: We observed a high specificity across all five brands of COVID-19 RDTs, ranging from 98.3 to 100%. Out of the confirmed COVID-19 samples, one patient tested positive for dengue IgM only, another tested positive for dengue IgG only. One patient tested positive for dengue IgG, IgM, and NS1, suggesting a co-infection. In COVID-19 IgG and/or IgM samples, 6.3% of COVID-19 IgG-positive samples also tested positive for dengue IgG, while 21.1% of COVID-19 IgM-positive samples also tested positive for dengue IgG.

Conclusion: Despite the high specificity of the COVID-19 RDT, we observed cross-reactions and false-positive results between dengue and COVID-19. Dengue and COVID-19 co-infection was also found. Health practitioners in dengue endemic areas should be careful when using antibody RDT for the diagnosis of dengue during the COVID-19 pandemic to avoid misdiagnosis.

Keyword: COVID-19, RDT, IgG, IgM, Dengue, Specificity, Cross-reactivity

Background

Severe acute respiratory coronavirus 2 (SARS-CoV-2) emerged in Wuhan, China, causing a respiratory disease, coronavirus disease 2019 (COVID-19), and has

now resulted in a global pandemic [1]. The pandemic remains ongoing in many countries, including areas where dengue is endemic, such as Indonesia, which adds a burden to health systems [2, 3]. There were over 130 000 reported cases of dengue in Indonesia in 2019 with an incidence rate of 51.48 cases per 100 000 population, an increase from the previous year's incidence of 24.75 cases per 100 000 population. As of 21 June, there are 68 000 cases of dengue reported across Indonesia in 2020, while COVID-19 cases continue to increase [4]. As of 26

*Correspondence: sasmono@ejikman.go.id

¹ Dengue Research Unit, Eijkman Institute for Molecular Biology, Ministry of Research and Technology/National Agency for Research and Innovation of the Republic of Indonesia, Jl. Diponegoro 69, Jakarta 10430, Indonesia

Full list of author information is available at the end of the article



© The Author(s) 2021. **Open Access** This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons licence, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons licence, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons licence and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this licence, visit <http://creativecommons.org/licenses/by/4.0/>. The Creative Commons Public Domain Dedication waiver (<http://creativecommons.org/publicdomain/zero/1.0/>) applies to the data made available in this article, unless otherwise stated in a credit line to the data.

Jan 2021, there are over one million confirmed cases of COVID-19 in Indonesia [5]. Dengue fever and COVID-19 have similar clinical and laboratory features, which can lead to misdiagnosis, delayed treatment, and isolation [3]. In both cases, patients often report acute fever, myalgia, fatigue, and other flu-like symptoms, as well as present with thrombocytopenia and leukopenia [1, 3]. Most commercial rapid diagnostic tests (RDT) available in the market are for the detection of SARS-CoV-2 antibodies, with relatively high sensitivity and specificity, especially when samples are taken later in the disease progression [6]. However, it is hampered by the apparent cross-reactivity resulting in false-positive results [7]. For dengue, immunochromatographic tests for the detection of dengue virus nonstructural protein 1 (NS1) antigen, IgM, IgG, and IgA antibodies have been developed by a number of commercial companies and have found wide application because of their ease of use and rapidity of results [8, 9].

Methods

In this study, we assessed the possibility of dengue and SARS-CoV-2 antibody cross-reactivity using three strategies. Firstly, we evaluate the specificity of five COVID-19 RDT brands against 60 well-characterized

RT-PCR-confirmed dengue patient's serum panel. Secondly, we test 95 RT-PCR-confirmed clinical COVID-19 samples on dengue RDT. And thirdly, we test 49 sera from healthy, asymptomatic individuals that are positives for COVID-19 IgG and/or IgM antibodies on dengue RDT (Table 1). The use of archived dengue patients' samples has been approved by Eijkman Institute Research Ethics Committee, approval number 151/2020, while the use of COVID-19 RT-PCR confirmed samples has been approved by Bali Mandara District Hospital Health Research Ethics Committee, approval number 007/EA/KEPK.RSBM.DISKES/2020, while the use of COVID-19 IgG and/or IgM positive samples has been approved by Raden Mattaher Hospital Research Ethics Committee, approval number S.32/SPE/VII/2020.

The specificity evaluation of COVID-19 IgG / IgM RDTs against confirmed dengue samples was performed in a laboratory that is internationally certified for Good Clinical Laboratory Practice (GCLP) at the Eijkman Institute, Jakarta, Indonesia. The laboratory testing on RT-PCR-confirmed COVID-19 samples was performed at Bali Mandara and Sanjiwani hospitals, Bali, Indonesia while the laboratory testing on healthy/asymptomatic samples positive for COVID-19 IgG and/or IgM was performed at Siloam Hospital, Jambi,

Table 1 Characteristics of samples used in the study

Categories	Dengue-confirmed samples (N = 60)	COVID-19-confirmed samples (N = 95)*	Healthy/asymptomatic samples (N = 49)
Fever day onset, mean (SD)	5 (1.5)	N/A	N/A
Age, median (IQR)	14 (8–22)	34 (25–46)	43 (34–52)
Age	N (%)		
Children ≤ 18 years	35 (58)	2 (2.1)	0 (0.0)
Adults > 18 years	25 (42)	78 (82.1)	49 (100)
Gender			
Male	28 (47)	49 (51.6)	34 (69.4)
Female	32 (53)	31 (32.6)	15 (30.6)
Serotype			11
DENV-1	15 (25)	N/A	N/A
DENV-2	15 (25)	N/A	N/A
DENV-3	20 (33)	2	N/A
DENV-4	10 (17)	N/A	N/A
Immunologic status			
Primary dengue infection	10 (17)	N/A	N/A
Secondary dengue infection	50 (83)	N/A	N/A
Presence of anti-dengue antibodies			
IgG (+), IgM (+)	21 (35)	N/A	N/A
IgG (+), IgM (–)	16 (27)	N/A	N/A
IgG (–), IgM (+)	14 (23)	N/A	N/A
IgG (–), IgM (–)	9 (15)	N/A	N/A

*Incomplete data for 15 samples

Indonesia, both with appropriate biosafety protocols. Dengue virus (DENV) RNA in the sera were extracted using QIAamp Viral RNA mini kit (Qiagen, Germany) according to the manufacturer's instructions. Detection and serotyping of DENV was performed using CDC DENV-1-4 real-time RT-PCR assay according to given instructions (Package Insert, 2018 available at www.cdc.gov/dengue). Molecular detection of SARS-CoV-2 was performed using A*Star Fortitude COVID-19 Real-Time RT-PCR Test and Liferiver Novel Coronavirus (2019-nCoV) Real Time Multiplex RT-PCR (BioVendor, Czech Republic). Immunologic status for dengue was determined using NovaLisa Dengue Virus IgG ELISA (NovaTec Immundiagnostica, Germany), performed according to manufacturer's instruction. The confirmed-dengue samples were collected from various regions across Indonesia between October 2018 and March 2019, before the initial outbreak of COVID-19. The confirmed-COVID-19 patients were recruited from Bali, Indonesia in May 2020. The healthy/asymptomatic individuals that were positive for COVID-19 IgG and/or IgM were recruited from Jambi, Indonesia from May to July 2020.

The COVID-19 IgG and IgM RDTs that we evaluated were Cellex Inc. (USA), Dynamiker Biotechnology (Tianjin, China), Genbody Inc. (Korea), Standard Diagnostics (Korea), and VivaDiag (Vivachek Biotech Hangzhou, China). Each dengue-confirmed sample was tested using different brands of IgG/IgM COVID-19 RDT simultaneously, under uniform conditions, according to their respective manufacturers' instructions. Positive COVID-19 RDT results were confirmed using SARS-CoV-2 Spike S1-RBD IgG & IgM ELISA Detection Kit (Genscript, USA). The dengue RDT kit used on confirmed COVID-19 samples and healthy/asymptomatic sample was Standard Q Dengue Duo (SD Biosensor, Korea, for the detection of dengue NS1 antigen and IgG and IgM antibodies), while the COVID-19 kit used on the healthy/asymptomatic samples are VivaDiag COVID-19 IgG/IgM.

For COVID-19 RDT specificity evaluation, which is the proportion of non-COVID-19 samples who are correctly identified as negative for SARS-CoV-2 antibodies using the COVID-19 RDTs, sample size was calculated based on the three-step method [10] with an expected specificity of 96%, a precision of 7.5%, and 90% power, which yields 60 presumed COVID-19-negative samples. Overall specificity of each RDT brand, as well as when stratified into variable groups, were compared using Z-tests for proportions. Specificity, confidence intervals, and other statistical analyses were performed using R Studio software, with *p*-value of less than 0.05 denoting statistical significance.

Results

We observed a high specificity across all COVID-19 RDT brands against dengue-confirmed samples, ranging from 98.3 to 100% and no statistically significant difference in overall performance between brands, or when grouped by age, gender, day of fever onset, DENV serotype, and immunologic status (Table 2). One sample which tested positive for COVID-19 IgG using SD Biosensor Dengue RDT was a 17-year-old male, with fever day onset of 2 days, infected with DENV-4, and tested positive for anti-dengue IgG and IgM. One other sample which tested positive for COVID-19 IgM using Cellex and SD Biosensor Dengue RDT was a 23-year-old male, with fever day onset of 2 days, infected with DENV-2, and tested positive for anti-dengue IgM and negative for anti-dengue IgG. However, neither of these samples tested positive for IgG or IgM using SARS-CoV-2 ELISA, suggesting false-positive results on the RDT.

Among 95 RT-PCR confirmed COVID-19 patient samples, one patient tested positive for dengue IgM only, another tested positive for dengue IgG only. One patient tested positive for both dengue IgG and IgM as well as NS1, demonstrating the possibility of coinfection of dengue and COVID-19 or recent dengue infection (Table 3).

In healthy/asymptomatic samples, out of 33 samples that tested positive for COVID-19 IgG, 2 samples (6.1%) also tested positive for dengue IgG, while out of 19 samples that tested positive for COVID-19 IgM, 4 samples (21.1%) also tested positive for dengue IgG (Table 3).

Discussion

While there is a possibility of cross-reactivity between SARS-CoV-2 and DENV antibodies, there is good overall performance of COVID-19 RDTs. Another study shows no cross reactivity with other respiratory

Table 2 Specificity comparison of five brands of COVID-19 RDTs against dengue-confirmed samples

	Specificity (%; 95% CI)	
	IgG	IgM
Cellex	100% (95% CI 92.5–100)	98.3% (95% CI 89.9–99.9)
Dynamiker	100% (95% CI 92.5–100)	100% (95% CI 92.5–100)
Genbody	100% (95% CI 92.5–100)	100% (95% CI 92.5–100)
Standard diagnostics	98.3% (95% CI 89.9–99.9)	98.3% (95% CI 89.9–99.9)
VivaChek	100% (95% CI 92.5–100)	100% (95% CI 92.5–100)
<i>p</i> -value	0.404	0.555

Table 3 Positivity of COVID-19-related samples on dengue NS1, IgG, and IgM RDT

Samples categories	N	Dengue RDT		
		NS1-positive N (%)	IgG-positive N (%)	IgM-positive N (%)
COVID-19 RT-PCR positive	95	1 (1.1)	2 (2.1)	2 (2.1)
COVID-19 IgG RDT positive	33	0 (0.0)	2 (6.1)	0 (0.0)
COVID-19 IgM RDT positive	19	0 (0.0)	4 (21.1)	0 (0.0)

pathogens that may have similar clinical symptoms as COVID-19 [11]. Apparent cross-reactivity between DENV and SARS-CoV-2 antibodies in dengue RDT has been reported in two confirmed COVID-19 cases in Singapore [7]. However, more in-depth studies are needed to fully understand the interaction between these antibodies.

We found that in terms of specificity, different COVID-19 RDT brands perform similarly to each other, comparable to findings in other studies [12]. However, these studies also show that different RDT brands have varying levels of sensitivity [12]. A meta-analysis of COVID-19 serological tests measured the pooled sensitivity of lateral flow immunoassays to be 66.0% (95% CI 49.3–79.3%) [13]. Taking that sensitivity as a reference, the rapid tests in our study have a negative predictive value (NPV) of 92.2% even at populations with 20% prevalence. The ability to rapidly confirm the absence of COVID-19 would be useful in allocating resources for isolation and protecting the patient from nosocomial COVID-19 infections especially in settings where the patient may have other viral infections that require hospitalization, such as dengue. Additionally, the ability to accurately rule out dengue in acutely febrile patients would quickly lead to the investigation and management of other differential diagnoses such as COVID-19 that would require timely isolation and contact tracing.

One of the limitations in this study is we did not assess the sensitivity of COVID-19 RDT data. Another limitation is that we could not conduct repeat testing of dengue antibodies in the COVID-19 confirmed samples and therefore cannot rule out the possibility of recent but not concurrent dengue infection. Additionally, conditions on site unfortunately did not make it possible to follow-up with RT-PCR for DENV. However, the positive anti dengue virus NS1 antigen result has been widely accepted as “confirmed dengue” (WHO SEARO Dengue Guidelines 2011). We also could not assess for cross-reactivity

with other coronavirus infections, auto-immune disease, or immunodeficiency. Nevertheless, our data provide important information on the prospective use of these RDTs in regions that are endemic for dengue during the COVID-19 pandemic.

Conclusion

Our study observed that despite the high specificity of the COVID-19 RDT, cross-reactions and false-positive results between dengue and COVID-19 are very likely. Dengue and COVID-19 co-infection can also occur. To the authors' knowledge, this is the first study reporting on the performance of COVID-19 rapid tests on dengue-confirmed samples. Our study adds important information on the possible cross-reaction between dengue and COVID-19 in Indonesia, though more studies are needed to further evaluate these interactions. Health practitioners should be careful when using antibody RDT for the diagnosis of dengue during the COVID-19 pandemic to avoid misdiagnosis.

Abbreviations

COVID-19: coronavirus disease 2019; DENV: dengue virus; ELISA: enzyme-linked immunosorbent assay; IgA: immunoglobulin A; IgG: immunoglobulin G; IgM: immunoglobulin M; NS1: nonstructural protein 1; RDT: rapid diagnostic test; RNA: ribonucleic acid; RT-PCR: reverse transcription polymerase chain reaction; SARS-CoV-2: severe acute respiratory coronavirus 2.

Acknowledgements

We would like to thank patients, nurses, and health practitioners involved in this study.

Authors' contributions

RTS and MLH set the conceptual design of the study; MSS, SM, and SH performed diagnosis, patients' recruitment, data collection, and analysis on patients' status. MSS and RTS made the first draft of the manuscript; BY, RTS and MLH contributed to the generation of final manuscript.

Funding

This work was supported by PT ASTRA International and the Ministry of Research and Technology / National Agency for Research and Innovation of the Republic Indonesia.

Availability of data and materials

RDT evaluation raw data available upon request.

Declarations

Ethics approval and consent to participate

The use of archived dengue patients' samples has been approved by Eijkman Institute Research Ethics Committee, approval number 151/2020, while the use of COVID-19 RT-PCR confirmed samples has been approved by Bali Mandara District Hospital Health Research Ethics Committee, approval number 007/EA/KEPKRSBM.DISKES/2020, while the use of COVID-19 IgG and/or IgM positive samples has been approved by Raden Mattaher Hospital Research Ethics Committee, approval number 5.32/SPE/VII/2020.

Consent for publication

N/A.

Competing interests

The authors declare that they have no competing interests. PT Kirana Jaya Lestari donated the VivaDiag RDT.

Author details

¹ Dengue Research Unit, Eijkman Institute for Molecular Biology, Ministry of Research and Technology/National Agency for Research and Innovation of the Republic Indonesia, Jl. Diponegoro 69, Jakarta 10430, Indonesia. ² Department of Internal Medicine, Faculty of Medicine and Health Sciences, Universitas Warmadewa, Denpasar, Bali 80235, Indonesia. ³ Raden Mattaher Hospital, Jambi 36361, Indonesia. ⁴ Department of Pathogen Molecular Biology, London School of Hygiene and Tropical Medicine, London WC1E 7HT, UK.

Received: 10 November 2020 Accepted: 2 March 2021

Published online: 11 March 2021

References

- Huang Y, Tu M, Wang S, Chen S, Zhou W, Chen D, et al. Clinical characteristics of laboratory confirmed positive cases of SARS-CoV-2 infection in Wuhan, China: a retrospective single center analysis. *Travel Med Infect Dis*. 2020. <https://doi.org/10.1016/j.tmaid.2020.101606>.
- Wu D, Lu J, Liu Q, Ma X, He W. To alert coinfection of COVID-19 and dengue virus in developing countries in the dengue-endemic area. *Infect Control Hosp Epidemiol*. 2020. <https://doi.org/10.1017/ice.2020.187>.
- Harapan H, Ryan M, Yohan B, Abidin RS, Nainu F, Rakib A, et al. Covid-19 and dengue: double punches for dengue-endemic countries in Asia. *Rev Med Virol*. 2020. <https://doi.org/10.1002/rmv.2161>.
- Huang E. Outbreak of dengue fever in Southeast Asia is “exploding” amid the coronavirus fight. *CNBC*. 2020. <https://www.cnbc.com/2020/07/10/outbreak-of-dengue-fever-in-southeast-asia-is-exploding-amid-the-coronavirus-fight.html>. Accessed 6 Aug 2020.
- COVID-19 GTPP. Berita Terkini | Gugus Tugas Percepatan Penanganan COVID-19. Covid19Gold n.d. <https://covid19.go.id/p/berita>. Accessed 2 Feb 2021.
- Castro R, Luz PM, Wakimoto MD, Veloso VG, Grinsztajn B, Perazzo H. COVID-19: a meta-analysis of diagnostic test accuracy of commercial assays registered in Brazil. *Braz J Infect Dis*. 2020;24:180–7. <https://doi.org/10.1016/j.bjid.2020.04.003>.
- Yan G, Lee CK, Lam LTM, Yan B, Chua YX, Lim AYN, et al. Covert COVID-19 and false-positive dengue serology in Singapore. *Lancet Infect Dis*. 2020;20:536. [https://doi.org/10.1016/S1473-3099\(20\)30158-4](https://doi.org/10.1016/S1473-3099(20)30158-4).
- Blacksell SD. Commercial dengue rapid diagnostic tests for point-of-care application: recent evaluations and future needs? *J Biomed Biotechnol*. 2012;2012:151967. <https://doi.org/10.1155/2012/151967>.
- Santoso MS, Yohan B, Denis D, Hayati RF, Haryanto S, Trianty L, et al. Diagnostic accuracy of 5 different brands of dengue virus non-structural protein 1 (NS1) antigen rapid diagnostic tests (RDT) in Indonesia. *Diagn Microbiol Infect Dis*. 2020;98:115116. <https://doi.org/10.1016/j.diagmicrobio.2020.115116>.
- Hess A, Shardell M, Johnson J, Thom K, Strassle P, Netzer G, et al. Methods and recommendations for evaluating and reporting a new diagnostic test. *Eur J Clin Microbiol Infect Dis*. 2012;31:2111–6. <https://doi.org/10.1007/s10096-012-1602-1>.
- Wu J-L, Tseng W-P, Lin C-H, Lee T-F, Chung M-Y, Huang C-H, et al. Four point-of-care lateral flow immunoassays for diagnosis of COVID-19 and for assessing dynamics of antibody responses to SARS-CoV-2. *J Infect*. 2020. <https://doi.org/10.1016/j.jinf.2020.06.023>.
- Van Elslande J, Houben E, Depypere M, Brackenier A, Desmet S, André E, et al. Diagnostic performance of seven rapid IgG/IgM antibody tests and the Euroimmun IgA/IgG ELISA in COVID-19 patients. *Clin Microbiol Infect*. 2020;26:1082–7. <https://doi.org/10.1016/j.cmi.2020.05.023>.
- Bastos ML, Tavaziva G, Abidi SK, Campbell JR, Haraoui L-P, Johnston JC, et al. Diagnostic accuracy of serological tests for covid-19: systematic review and meta-analysis. *BMJ*. 2020. <https://doi.org/10.1136/bmj.m2516>.

22 blisher's Note

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

Ready to submit your research? Choose BMC and benefit from:

- fast, convenient online submission
- thorough peer review by experienced researchers in your field
- rapid publication on acceptance
- support for research data, including large and complex data types
- gold Open Access which fosters wider collaboration and increased citations
- maximum visibility for your research: over 100M website views per year

At BMC, research is always in progress.

Learn more biomedcentral.com/submissions



Assessment of dengue and COVID-19 antibody rapid diagnostic tests cross-reactivity in Indonesia

ORIGINALITY REPORT

23%
SIMILARITY INDEX

16%
INTERNET SOURCES

19%
PUBLICATIONS

7%
STUDENT PAPERS

PRIMARY SOURCES

- 1

Marsha S. Santoso, Benediktus Yohan, Dionisius Denis, Rahma F. Hayati et al. "Diagnostic accuracy of 5 different brands of dengue virus non-structural protein 1 (NS1) antigen rapid diagnostic tests (RDT) in Indonesia", Diagnostic Microbiology and Infectious Disease, 2020
Publication

3%
- 2

www.medrxiv.org
Internet Source

2%
- 3

Boon-Teong Teoh, Sing-Sin Sam, Kim-Kee Tan, Jefree Johari, Juraina Abd-Jamil, Poh-Sim Hooi, Sazaly AbuBakar. "The Use of NS1 Rapid Diagnostic Test and qRT-PCR to Complement IgM ELISA for Improved Dengue Diagnosis from Single Specimen", Scientific Reports, 2016
Publication

2%
- 4

www.pointstyle.fr
Internet Source

1%

5	www.biomedcentral.com Internet Source	1 %
6	www.researchsquare.com Internet Source	1 %
7	onlinelibrary.wiley.com Internet Source	1 %
8	Tsai-Ying Yen, Maria de Jesus Trovoada dos Santos, Lien-Feng Tseng, Shu-Feng Chang et al. "Seroprevalence of antibodies against dengue virus among pregnant women in the Democratic Republic of Sao Tome and Principe", Acta Tropica, 2016 Publication	1 %
9	www.cdc.gov Internet Source	1 %
10	journals.plos.org Internet Source	1 %
11	www.abs.gov.au Internet Source	1 %
12	link.springer.com Internet Source	1 %
13	Aisling M Vaughan, Erika F Duffell, Gudrun S Friedl, Diogo S Lemos et al. "Seroprevalence of SARS-CoV-2 antibodies prior to the widespread introduction of vaccine	1 %

programmes in the WHO European Region,
January - December 2020: a systematic
review", Cold Spring Harbor Laboratory, 2021

Publication

14	www.jmb.or.kr	1 %
----	--	-----

Internet Source

15	wwwnc.cdc.gov	1 %
----	--	-----

Internet Source

16	R. Tedjo Sasmono, Angeline Sutjianto, Marsha S. Santoso, Karomah Sriwedari et al.	<1 %
----	---	------

"Molecular epidemiology of dengue in North Kalimantan, a province with the highest incidence rates in Indonesia in 2019",
Infection, Genetics and Evolution, 2021

Publication

17	Di Wu, Jianyun Lu, Qun Liu, Xiaowei Ma, Weiyun He. "To alert co-infection of SARS-COV-2 and dengue virus in developing countries in the dengue-endemic area",	<1 %
----	---	------

Infection Control & Hospital Epidemiology, 2020

Publication

18	SZU.CZ	<1 %
----	------------------------------------	------

Internet Source

19	www.fortunejournals.com	<1 %
----	--	------

Internet Source

Www.dovepress.com

20

Internet Source

<1 %

21

bjgpopen.org

Internet Source

<1 %

22

www.springermedizin.at

Internet Source

<1 %

23

assets.researchsquare.com

Internet Source

<1 %

24

index.mirasmart.com

Internet Source

<1 %

25

arupconsult.com

Internet Source

<1 %

26

eprints.lancs.ac.uk

Internet Source

<1 %

27

www.dovepress.com

Internet Source

<1 %

28

www.tandfonline.com

Internet Source

<1 %

29

Marsha S. Santoso, Benediktus Yohan, Dionisius Denis, Rahma F. Hayati et al. "Diagnostic accuracy of five different Brands of Dengue Virus non-Structural Protein 1 (NS1) antigen rapid diagnostic tests (RDT) in Indonesia", Diagnostic Microbiology and Infectious Disease, 2020

<1 %

30

Ruxandra Mare, Ioan Sporea, Raluca Lupușoru, Roxana Șirli, Alina Popescu, Mirela Danila, Corina Pienar. "The value of ElastPQ for the evaluation of liver stiffness in patients with B and C chronic hepatopathies", *Ultrasonics*, 2017

Publication

<1 %

31

repositorio.ipen.br

Internet Source

<1 %

32

Victoria Higgins, Anselmo Fabros, Xiao Yan Wang, Maria Bhandari, David J. Daghfal, Vathany Kulasingam. "Anti-SARS-CoV-2 IgM improves clinical sensitivity early in disease course", *Clinical Biochemistry*, 2021

Publication

<1 %

33

Aryati Aryati, Hidayat Trimarsanto, Benediktus Yohan, Puspa Wardhani, Sukmal Fahri, R Tedjo Sasmono. "Performance of commercial dengue NS1 ELISA and molecular analysis of NS1 gene of dengue viruses obtained during surveillance in Indonesia", *BMC Infectious Diseases*, 2013

Publication

<1 %

34

F. Jurgens, B. Hogema, S. Siegerink, L. Samwel et al. "The Persistence of Natural SARS-CoV-2 Antibodies 12 Months Post COVID-19, a

<1 %

Prospective Cohort Study.", Research Square Platform LLC, 2021

Publication

35

Prasetyo Hariadi, Dewi Lokida, Adhella M Naysilla, Nurhayati Lukman et al. "Coinfection with SARS-CoV-2 and dengue virus: a case report", Research Square, 2020

Publication

<1 %

36

Xingjian Hu, Yin Wang, Junwei Liu, Xuefeng Qiu et al. "Early outcomes of Stanford type A aortic dissection under the coronavirus disease 2019 (COVID-19) pandemic: a multicentre study from Hubei province", Interactive CardioVascular and Thoracic Surgery, 2020

Publication

<1 %

Exclude quotes On

Exclude matches Off

Exclude bibliography On